

PRODUCT CHANGE NOTIFICATION

PCN #: PCN-18-1228	INITIATOR: Davis Wang	Issue Date: 12/28/2018
---------------------------	------------------------------	-------------------------------

PRODUCTS AFFECTED

Customer Part No.	Inventus Part No.	Description
MWA180024A-12A, MWA220024A-12A,	03-57616-624-03 03-57616-624-01	IR POWER SUPPLY MEDICAL180W 24V IR POWER SUPPLY MEDICAL220W 24V

CHANGE INFORMATION

Reason for Change:

1. Apply for new agency certification

Description of Change:

1. add PSE logo on unit label per agency request, others keep same as before
2. Delete capacitor CY6 & CY7(30LVT22-R& CS45-B2GA221K-VKA),
3. Add resistor R98/R99/R100(330kΩ 0.25W), PCBA re-layout
4. Change fuse to 215 series (250Vac, T4AH, 5X20mm)

The label drawing updated as below.



90939-894 C.pdf 90939-812 C.pdf

Qualification Results:

OK

Marking Information to denote updated parts:

Inform customer

Implementation Date:

2 weeks after get customer feedback

Comments:

APPROVALS

Check this box if customer approval is not required. Please provide justification.

Justification: _____

Inventus Power	Program Manager	_____	_____	_____
		Name	Signature	Date
	Engineering		Davis	12/28/2018
		Name	Signature	Date
Customer	Product Assurance	_____	_____	_____
		Name	Signature	Date
	Customer		YO	12/28/2018
		Name	Signature	Date
	Customer	_____	_____	_____
		Name	Signature	Date